

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/22/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:58
INTERNAL STANDARD	VERIFIED	06:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:59
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:00
BLANK TEST	0.00	07:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:01
BLANK TEST	0.00	07:02

J. M. Fish #27

