

SOLUTION CHANGE

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 10/01/2024

OPERATOR NAME:
 STEVEN T BUNIO
 Permit #: 44983
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:31
INTERNAL STANDARD	VERIFIED	14:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:32
BLANK TEST	0.00	14:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:34
BLANK TEST	0.00	14:35

AS Shew

