

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 10/22/2024

OPERATOR NAME:
 J RIEHL
 Permit #: 47503
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:28
INTERNAL STANDARD	VERIFIED	03:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:29
BLANK TEST	0.00	03:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:30
BLANK TEST	0.00	03:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:31
BLANK TEST	0.00	03:32

[Handwritten signature]
 #870

