

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/15/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:15
INTERNAL STANDARD	VERIFIED	07:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:17
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:18
BLANK TEST	0.00	07:19

POST #427

