

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/22/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:34
INTERNAL STANDARD	VERIFIED	14:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:34
BLANK TEST	0.00	14:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:35
BLANK TEST	0.00	14:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:37
BLANK TEST	0.00	14:37

