

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 11/29/2024

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:06
INTERNAL STANDARD	VERIFIED	10:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:06
BLANK TEST	0.00	10:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:08
BLANK TEST	0.00	10:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:09
BLANK TEST	0.00	10:10

*PO Steve Bunio #789*

