

SOLUTION CHANGE




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 12/01/2024

OPERATOR NAME:
 JOHN D BATEHOLTS
 Permit #: 43007
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:39
INTERNAL STANDARD	VERIFIED	09:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:41
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:42
BLANK TEST	0.00	09:43



Handwritten signature and number 8714