

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:38
INTERNAL STANDARD	VERIFIED	21:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:41
BLANK TEST	0.00	21:42

#820

