

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 10/15/2024

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:03
INTERNAL STANDARD	VERIFIED	15:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:05
BLANK TEST	0.00	15:06

Signature #787

