

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 10/23/2024

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:30
BLANK TEST	0.00	10:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:32
BLANK TEST	0.00	10:33



*Charles B. Geisler*