

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 10/29/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:05
BLANK TEST	0.00	07:05

*PO Fish #427*

