

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 11/09/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:41
BLANK TEST	0.00	07:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44

*[Handwritten signature]*  
#800

