

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 11/15/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	07:18
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20

PO FISH #427

