

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 11/22/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:06
INTERNAL STANDARD	VERIFIED	10:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:06
BLANK TEST	0.00	10:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:07
BLANK TEST	0.00	10:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:09
BLANK TEST	0.00	10:09

Donnie J Long

