

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 12/08/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:42
INTERNAL STANDARD	VERIFIED	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:42
BLANK TEST	0.00	21:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:44
BLANK TEST	0.00	21:45

*[Handwritten signature]*  
#820

