

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 12/16/2024

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:09
INTERNAL STANDARD	VERIFIED	10:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	10:11
BLANK TEST	0.00	10:12



SGT *[Signature]* #157