

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113608

Date: 12/23/2024

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:05
INTERNAL STANDARD	VERIFIED	08:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	08:05
BLANK TEST	0.00	08:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:07
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:08
BLANK TEST	0.00	08:08



Set *[Signature]* #152