

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113608

Date: 12/27/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:11
INTERNAL STANDARD	VERIFIED	07:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:13
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:14

*RF #427*

