

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/08/2024

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:15
INTERNAL STANDARD	VERIFIED	11:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:15
BLANK TEST	0.00	11:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:16
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:17
BLANK TEST	0.00	11:18

rc

