

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 01/22/2024

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:11
INTERNAL STANDARD	VERIFIED	17:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:12
BLANK TEST	0.00	17:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:13
BLANK TEST	0.00	17:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:14
BLANK TEST	0.00	17:15

*PO Steven Bunio #789*

