

SOLUTION CHANGE

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 02/01/2024

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:32
INTERNAL STANDARD	VERIFIED	09:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:32
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:33
BLANK TEST	0.00	09:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:34
BLANK TEST	0.00	09:35

