

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:31
BLANK TEST	0.00	09:32

[Handwritten signature]
#820

