

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/22/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:54
INTERNAL STANDARD	VERIFIED	08:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:55
BLANK TEST	0.00	08:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:57
BLANK TEST	0.00	08:58

Donnie J Long 844

