

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/02/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:38
INTERNAL STANDARD	VERIFIED	09:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:40
BLANK TEST	0.00	09:41

Don J Long
844

