

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/08/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:12
INTERNAL STANDARD	VERIFIED	16:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:13
BLANK TEST	0.00	16:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:14
BLANK TEST	0.00	16:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:15
BLANK TEST	0.00	16:16

PO [Signature] #4027

