

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/15/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:45
INTERNAL STANDARD	VERIFIED	09:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:45
BLANK TEST	0.00	09:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:47
BLANK TEST	0.00	09:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:48
BLANK TEST	0.00	09:49

[Handwritten signature]
#820

