

SUPERVISOR TEST


WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:33
INTERNAL STANDARD	VERIFIED	14:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:34
BLANK TEST	0.00	14:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:35
BLANK TEST	0.00	14:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:36
BLANK TEST	0.00	14:37


#1820

