

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/22/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:49
INTERNAL STANDARD	VERIFIED	12:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:49
BLANK TEST	0.00	12:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:52
BLANK TEST	0.00	12:53

*[Handwritten signature]*  
#820

