

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/08/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:23
INTERNAL STANDARD	VERIFIED	13:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:23
BLANK TEST	0.00	13:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:25
BLANK TEST	0.00	13:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:26
BLANK TEST	0.00	13:27

PO JMF #4127

