

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/22/2024

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:38
INTERNAL STANDARD	VERIFIED	09:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:40
BLANK TEST	0.00	09:41



John D. Bateholts # 714

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506


Date: 05/22/2024

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:03
INTERNAL STANDARD	VERIFIED	08:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD INTERFERENCE		08:04

Interference Detected



John D. Bateholts 11717

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506


Date: 05/22/2024

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:07
INTERNAL STANDARD	VERIFIED	08:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD INTERFERENCE		08:07

Interference Detected



P. [Signature] 1714

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506


Date: 05/22/2024

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:24
INTERNAL STANDARD	VERIFIED	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD INTERFERENCE		08:25

Interference Detected



Handwritten signature and number: [Signature] 0714