

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/22/2024

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:51
INTERNAL STANDARD	VERIFIED	23:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:52
BLANK TEST	0.00	23:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:53
BLANK TEST	0.00	23:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:54
BLANK TEST	0.00	23:55



SGT [Signature] #757