

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/28/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:02
INTERNAL STANDARD	VERIFIED	15:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:05
BLANK TEST	0.00	15:06

