

SOLUTION CHANGE

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/22/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:11
INTERNAL STANDARD	VERIFIED	18:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:12
BLANK TEST	0.00	18:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:14
BLANK TEST	0.00	18:15

*PO FISH #427*

