

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/29/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:27
INTERNAL STANDARD	VERIFIED	10:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:28
BLANK TEST	0.00	10:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:29
BLANK TEST	0.00	10:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:30
BLANK TEST	0.00	10:31

[Handwritten signature]
#870

