

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:47
INTERNAL STANDARD	VERIFIED	11:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:48
BLANK TEST	0.00	11:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:49
BLANK TEST	0.00	11:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:50
BLANK TEST	0.00	11:51

[Handwritten signature]
#820

