

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/15/2024

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:55
INTERNAL STANDARD	VERIFIED	08:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:57
BLANK TEST	0.00	08:58



Handwritten signature and number: J. B. 2714