

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/22/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:28
INTERNAL STANDARD	VERIFIED	10:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:28
BLANK TEST	0.00	10:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:30
BLANK TEST	0.00	10:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:31
BLANK TEST	0.00	10:32


#820

