

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/29/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:48
INTERNAL STANDARD	VERIFIED	10:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:49
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:50
BLANK TEST	0.00	10:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:51
BLANK TEST	0.00	10:52

PO Fish #421

