

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/10/2024

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:14
INTERNAL STANDARD	VERIFIED	10:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:16
BLANK TEST	0.00	10:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:17
BLANK TEST	0.00	10:18

*Donnie J Long*

