

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/22/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:30
INTERNAL STANDARD	VERIFIED	15:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:31
BLANK TEST	0.00	15:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:32
BLANK TEST	0.00	15:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:33
BLANK TEST	0.00	15:34

