

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/01/2024

OPERATOR NAME:
S T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:42
INTERNAL STANDARD	VERIFIED	09:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:43
BLANK TEST	0.00	09:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:44
BLANK TEST	0.00	09:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:45
BLANK TEST	0.00	09:46

PO Shun 1/89

