

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/08/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:44
INTERNAL STANDARD	VERIFIED	14:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:45
BLANK TEST	0.00	14:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:46
BLANK TEST	0.00	14:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:47
BLANK TEST	0.00	14:48

