

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/15/2024

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:15
INTERNAL STANDARD	VERIFIED	10:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:15
BLANK TEST	0.00	10:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:16
BLANK TEST	0.00	10:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:17
BLANK TEST	0.00	10:18

