

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/22/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:11
INTERNAL STANDARD	VERIFIED	02:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:11
BLANK TEST	0.00	02:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:15

[Handwritten signature]
#820

