

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 10/29/2024

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:01
INTERNAL STANDARD	VERIFIED	15:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:01
BLANK TEST	0.00	15:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:04
BLANK TEST	0.00	15:05

*PO JMF #427*

