

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/01/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:18
INTERNAL STANDARD	VERIFIED	23:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:19
BLANK TEST	0.00	23:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:20
BLANK TEST	0.00	23:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22

