

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 11/08/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:13
INTERNAL STANDARD	VERIFIED	22:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:15
BLANK TEST	0.00	22:16

*[Handwritten signature]*  
#520

