

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/22/2024

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:47
INTERNAL STANDARD	VERIFIED	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:49
BLANK TEST	0.00	08:50

Don J
844

