

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/08/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:49
INTERNAL STANDARD	VERIFIED	22:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:49
BLANK TEST	0.00	22:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:51
BLANK TEST	0.00	22:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53

J Riehl
#1870

