

SUPERVISOR TEST



WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/16/2024

OPERATOR NAME:
BRANDON D KINDERMAN
Permit #: 38773
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:35
INTERNAL STANDARD	VERIFIED	16:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:35
BLANK TEST	0.00	16:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:36
BLANK TEST	0.00	16:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:38
BLANK TEST	0.00	16:38



J.P. [Signature] #126

SGT [Signature] #257